



CSA INTERNATIONAL

# Certificate of Compliance

**Certificate:** 1167960 (LR113484-3)

**Master Contract:** 176045

**Project:** 1557172

**Date Issued:** 2004/05/19

**Issued to:** Carl Zeiss Jena GmbH  
Pruefzentrum  
Carl-Zeiss-Promenade 10  
Jena, 07745  
Germany  
Attention: Dr. Wolfram Daenhardt

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'*



**Issued by:** Peter Cheung, P. Eng.

**Authorized by:** Gianluca Arcari, P.Eng.,  
Product Group Manager

## **PRODUCTS**

**CLASS 8721 84** - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

**CLASS 8721 04** - LABORATORY EQUIPMENT - Electrical

Microscopes Axiostar and Axiostar Plus, identification numbers 1122-100, 1122-099, 1122-098, 1169-149, 1169-150, 1169-151, cord connected, rated 100-240 V, 50/60 Hz, 65 VA, Pollution Degree 2, Installation Category II, IP20.

## **APPLICABLE REQUIREMENTS**

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



**Certificate:** 1167960 (LR113484-3)

**Master Contract:** 176045

**Project:** 1557172

**Date Issued:** 2004/05/19

---

CAN/CSA-C22.2 No. 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 1010.1B-92 - Amendment 2: 1997 to CAN/CSA-C22.2 No. 101.1-92, Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std No. 61010A-1 - Electrical Equipment for Laboratory Use; Part 1: General Requirements